



Dissolution Cable Kit

Part number: 9910097900

Installation category I
Pollution degree 2
Safety class 3 (EN61010-1)

Introduction

The Dissolution Cable kit is used to interface the 18 Cell Holder on the Cary, and also to provide a connection between the peristaltic pump and the Cary.

The instructions for installing the accessory are provided online with the Cary WinUV software. To view these instructions, start the Cary WinUV software, choose 'Help Topics' from the 'Help' menu and click 'Accessories' on the Cary WinUV Help home page. Click the 'Dissolution Cable kit' option from the list of Cary 50 accessories.

Safety information

Before using this accessory, you must read the Safety Practices and Hazards section in your Cary operating instructions.

If the equipment is used in a manner not specified by the manufacturer, the protection provided by the equipment may be impaired.

Specifications

For environmental and power supply requirements, refer to your Cary operating instructions.

Connections

Computer internal: 40-way flat ribbon cable with 37-pin D-range connector (plugs into Cary 50 PC card)

Cary-computer cable: 37-pin D-range connector from computer and 25-pin D-range connector mounted inside sample compartment

Cary-pump connection: 25-pin connector for 5-pin DIN cable from peristaltic pump

Weight

Packed:	1.0 kg (2.2 lb)
Unpacked:	0.82 kg (1.8 lb)



Agilent Technologies

Dissolution Cable Kit

Dimensions (W x D x H)

Packed: 350 x 280 x 60 mm (15.2 x 11.4 x 2.7 in)
Cable length: 2 m (78.8 in)

This information is subject to change without notice.



8510165300

Part Number: 8510165300

Edition 01/11

Issue 2

Printed in Australia

© Agilent Technologies, Inc. 1999, 2011

Agilent Technologies

679 Springvale Road

Mulgrave, VIC 3170