

Ordering Numbers of the 930 / 940 Instrument Series

2.	X	X	X.	X	X	X	X
	930	940	942 ^{a)}	1 st Digit	2 nd Digit	3 rd Digit	4 th Digit
Main Unit			–	0			
	without oven	ONE	ONE	1			
			ONE/DeG	1	0	6	
	with oven	TWO		2			
Suppression			–		0		
	non	non			1		
	ChS	ChS			2		
	ChS/PP	ChS/PP	ChS/PP		3		
	SeS	SeS			4		
	SeS/PP	SeS/PP	SeS/PP		5		
Options	–	–				0	0
		Prep 1*				1	0
		Prep 2*	Prep 2			2	0
		Prep 3**				3	0
		HPG***	HPG			4	0
		LPG***				5	0
	DeG					6	0
			LQH			7	0

Example: 2.940.1530 940 Professional IC Vario ONE/SeS/PP/Prep 3

Legend

ONE	1 x IC Pump, Injection Valve, Degasser
TWO	2 x IC Pump, Injection Valve, Degasser
non	non suppressed
ChS	Chemical Suppression
ChS/PP	Chemical Suppression with Peristaltic Pump
SeS	Sequential Suppression
SeS/PP	Sequential Suppression with Peristaltic Pump
Prep 1	Peristaltic Pump
Prep 2	Peristaltic Pump with Injection Valve
Prep 3	Peristaltic Pump with SPM
HPG	High Pressure Gradient
LPG	Low Pressure Gradient
DeG	Degasser for Eluent and Sample
LQH	10-Port Selection Valve, Injection Valve, Peristaltic Pump

- a) only one functionality
 * only in combination with ONE/ non-suppressed, SeS, SeS/PP
 ** only in combination with ONE/SeS, SeS/PP
 *** only in combination with all ONE versions

Instruments:

Old Generation			New Generation	
881 / 882	Compact IC pro / plus	→	930	Compact IC Flex
850	Professional IC	→	940	Professional IC Vario
2.849.1030	Inline Eluent Preparation	→	941	Eluent Production Module
872	Extension Module	→	942	Extension Module Vario
886	Professional Thermostat / Reactor	→	943	Professional Thermostat / Reactor Vario
887	Professional UV/VIS Detector	→	944	Professional UV/VIS Detector Vario
896	Professional Detector	→	945	Professional Detector Vario

www.metrohm.com

