

941 Eluent Production Module



Manual

8.941.8001EN / 2020-01-08



Metrohm AG

CH-9100 Herisau

Switzerland

Phone +41 71 353 85 85

Fax +41 71 353 89 01

info@metrohm.com

www.metrohm.com

941 Eluent Production Module

Manual

Technical Communication
Metrohm AG
CH-9100 Herisau
techcom@metrohm.com

This documentation is protected by copyright. All rights reserved.

This documentation has been prepared with great care. However, errors can never be entirely ruled out. Please send comments regarding possible errors to the address above.

Table of contents

1	Introduction	1
1.1	Instrument description	1
1.1.1	Intended use	1
1.2	About the documentation	2
1.2.1	Symbols and conventions	2
1.3	Safety instructions	3
1.3.1	General notes on safety	3
1.3.2	Electrical safety	3
1.3.3	Tubing and capillary connections	4
1.3.4	Flammable solvents and chemicals	4
1.3.5	Recycling and disposal	4
2	Overview of the instrument	6
2.1	Front	6
2.2	Rear	7
3	System overview	8
4	Installation	10
4.1	Setting up the instrument	10
4.1.1	Packaging	10
4.1.2	Checks	10
4.1.3	Location	10
4.2	Base tray and bottle holder	10
4.2.1	Basic information on base tray and bottle holder	10
4.2.2	Mounting base tray and bottle holder (optional)	11
4.3	Connecting the eluent bottle	16
4.4	Connecting the concentrate bottle	19
4.5	Connecting the ultrapure water container	22
4.5.1	Connecting the ultrapure water canister	22
4.5.2	Connecting an ELGA PURELAB flex	25
4.6	Installing the Dosino	26
4.7	Connecting the instrument	29
4.7.1	Connecting the instrument to the ion chromatograph	29
4.7.2	Connecting the instrument to the power grid	29
5	Operation	31
6	Operation and maintenance	32
6.1	Care	32



6.2	Servicing the door	32
6.3	Dosino	33
6.4	Dosing unit	33
7	Troubleshooting	34
7.1	Dosino	34
7.2	Dosing unit	34
8	Technical specifications	35
8.1	Reference conditions	35
8.2	Ambient temperature	35
8.3	Level sensor	35
8.4	Housing	36
8.5	Weight	36
8.6	Dosino	36
8.7	Dosing unit	36
8.8	Power connection	36
8.9	Interfaces	37
9	Accessories	38
	Index	39

Table of figures

Figure 1	Front	6
Figure 2	Rear	7
Figure 3	Removing the base tray	12
Figure 4	Mounting the base tray	13
Figure 5	Removing the bottle holder	14
Figure 6	Mounting the bottle holder	15
Figure 7	Bottle cap for level sensor - Inserting the measuring sensor	16
Figure 8	Installing the eluent bottle cap with level sensor	17
Figure 9	Installing the concentrate bottle cap without level sensor	19
Figure 10	Installing the concentrate bottle cap with level sensor	20
Figure 11	Installing the ultrapure water canister cap without level sensor	22
Figure 12	Installing the ultrapure water canister cap with level sensor	23
Figure 13	Installing an ELGA PURELAB flex	25
Figure 14	Connecting the Dosino	27

1 Introduction

1.1 Instrument description

The **941 Eluent Production Module** allows you to produce eluent for your IC system from a concentrate and ultrapure water as needed and in a software-controlled manner.

Thanks to the built-in Level Control, the instrument is capable of monitoring the liquid levels in up to 4 liquid containers, if they are equipped with level sensors.

A level sensor for the eluent bottle is indispensable for the automated production of eluent and is contained in the scope of delivery of the 941 Eluent Production Module. However, we recommend using level sensors to also monitor the bottle with the concentrate and the ultrapure water container. Additional level sensors are available as optional accessories in various versions.

The 941 Eluent Production Module is operated with the **MagIC Net** software. If the 941 Eluent Production Module is connected to a 940 Professional IC Vario instrument, MagIC Net automatically detects the 941 Eluent Production Module and checks its functional capability. It controls and monitors the entire IC system including all connected instruments; it evaluates the measured data and manages it in a database.

1.1.1 Intended use

The 941 Eluent Production Module is an instrument for the automated production of eluent for ion chromatography systems. It can monitor the levels of various liquid containers and, where necessary, perform appropriate actions controlled by the software.

The present instrument is suitable for processing chemicals. Usage of the 941 Eluent Production Module therefore requires the user to have basic knowledge and experience in handling toxic and caustic substances. Knowledge regarding the application of fire prevention measures prescribed for laboratories is also mandatory.



1.2 About the documentation



CAUTION

Read through this documentation carefully before putting the instrument into operation. The documentation contains information and warnings which the user must follow in order to ensure safe operation of the instrument.

1.2.1 Symbols and conventions

The following symbols and formatting may appear in this documentation:

(5-12)

Cross-reference to figure legend

The first number refers to the figure number, the second to the instrument part in the figure.

1

Instruction step

Carry out these steps in the sequence shown.

Method

Dialog text, parameter in the software

File ▶ New

Menu or menu item

[Next]

Button or **key**



WARNING

This symbol draws attention to a possible life-threatening hazard or risk of injury.



WARNING

This symbol draws attention to a possible hazard due to electrical current.



WARNING

This symbol draws attention to a possible hazard due to heat or hot instrument parts.



WARNING

This symbol draws attention to a possible biological hazard.



CAUTION

This symbol draws attention to possible damage to instruments or instrument parts.

**NOTE**

This symbol highlights additional information and tips.

1.3 Safety instructions

1.3.1 General notes on safety

**WARNING**

Operate this instrument only according to the information contained in this documentation.

This instrument left the factory in a flawless state in terms of technical safety. To maintain this state and ensure non-hazardous operation of the instrument, the following instructions must be observed carefully.

1.3.2 Electrical safety

The electrical safety when working with the instrument is ensured as part of the international standard IEC 61010.

**WARNING**

Only personnel qualified by Metrohm are authorized to carry out service work on electronic components.

**WARNING**

Never open the housing of the instrument. The instrument could be damaged by this. There is also a risk of serious injury if live components are touched.

There are no parts inside the housing which can be serviced or replaced by the user.

Supply voltage

**WARNING**

An incorrect supply voltage can damage the instrument.

Only operate this instrument with a supply voltage specified for it (see rear panel of the instrument).



Protection against electrostatic charges



WARNING

Electronic components are sensitive to electrostatic charges and can be destroyed by discharges.

Do not fail to pull the power cord out of the power socket before you set up or disconnect electrical plug connections at the rear of the instrument.

1.3.3 Tubing and capillary connections



CAUTION

Leaks in tubing and capillary connections are a safety risk. Tighten all connections well by hand. Avoid applying excessive force to tubing connections. Damaged tubing ends lead to leakage. Appropriate tools can be used to loosen connections.

Check the connections regularly for leakage. If the instrument is used mainly in unattended operation, then weekly inspections are mandatory.

1.3.4 Flammable solvents and chemicals




WARNING

All relevant safety measures are to be observed when working with flammable solvents and chemicals.

- Set up the instrument in a well-ventilated location (e.g. fume cupboard).
- Keep all sources of flame far from the workplace.
- Clean up spilled liquids and solids immediately.
- Follow the safety instructions of the chemical manufacturer.

1.3.5 Recycling and disposal



This product is covered by European Directive 2012/19/EU, WEEE – Waste Electrical and Electronic Equipment.

The correct disposal of your old instrument will help to prevent negative effects on the environment and public health.



More details about the disposal of your old instrument can be obtained from your local authorities, from waste disposal companies or from your local dealer.



2 Overview of the instrument

2.1 Front




Figure 1 Front

1 Dosino holder

Receptacle for a Dosino.

2 LED

Standby indicator.

3 Cable duct

Conduit to the rear of the instrument. For feeding through tubing and the Dosino cable.

2.2 Rear




Figure 2 Rear

1 USB sockets

2 sockets for plugging in USB cables for connecting devices with USB connector. Labeled *USB 1* and *USB 2*.

3 MSB sockets

3 sockets for plugging in MSB cables for connecting devices with MSB connector. Labeled *MSB 2*, *MSB 3* and *MSB 4*.

5 Level sensor sockets

4 sockets for connecting level sensor cables (6.2151.060).

7 Cable duct

Conduit to the front of the instrument. For feeding through tubing and the Dosino cable.

2 Controller socket

Socket for plugging in the USB cable (6.2151.000) for connecting the ion chromatograph.


4 Power socket

For connecting the power cord.

6 Type plate

3 System overview

The following figure shows an ion chromatography system with a Sample Processor (left), the ion chromatograph (middle) and the 941 Eluent Production Module (right).



The 941 Eluent Production Module consists of a housing with a built-in Level Control. A Dosino with a 50 mL dosing unit, the canister for ultrapure water, the bottle for the eluent concentrate, the level sensor for the eluent bottle and all small parts required for a standard setup are included in the accessories supplied.

In a standard setup, only the eluent bottle is equipped with a level sensor. The ultrapure water canister and the bottle with the concentrate are not monitored.

If you intend to let your ion chromatography system run unsupervised for extended periods of time, we recommend also monitoring the concentrate bottle and the ultrapure water canister with a level sensor.

Level sensors come with measuring sensors in different lengths:

- Sensors with short measuring sensors warn when the liquid container is full (FULL sensor).
- Sensors with long measuring sensors warn when the liquid container is empty (EMPTY sensor). Sensors with measuring sensors of different lengths are available for 1 L bottles, 2 L bottles and for 10 L canisters.



4 Installation

4.1 Setting up the instrument

4.1.1 Packaging

The instrument is supplied in protective packaging together with the separately packed accessories. Keep this packaging, as only this ensures safe transportation of the instrument.

4.1.2 Checks

Immediately after receipt, check whether the shipment has arrived complete and without damage by comparing it with the delivery note.

4.1.3 Location

The 941 Eluent Production Module can be placed directly onto the 940 Professional IC Vario. To accomplish this, the bottle holder of the ion chromatograph has to be removed and placed onto the 941 Eluent Production Module (*see chapter 4.2.2.2, page 13*).

The 941 Eluent Production Module can also be placed directly under the 940 Professional IC Vario. To accomplish this, the base tray of the ion chromatograph has to be removed and installed again under the 941 Eluent Production Module (*see chapter 4.2.2.1, page 11*).

As an alternative, the 941 Eluent Production Module can be set up together with other instruments with the same footprint in a separate stack next to the ion chromatograph. We recommend that a bottle holder (6.2061.100) and a base tray (6.2061.110) be mounted for each stack of instruments.

4.2 Base tray and bottle holder

4.2.1 Basic information on base tray and bottle holder

The base tray (6.2061.110) and bottle holder (6.2061.100) protect IC instruments from dust, dirt and leaking fluids. The supply bottles for eluent and auxiliary solutions can be positioned neatly on the bottle holder.

In a complex IC system, several different instruments may be used, such as an analyzer, an extension module and a detector. These instruments can be set up in one or more stacks. We recommend that a base tray and bottle holder be mounted for each stack of IC instruments.

The bottle holder and base tray must be removed or set up if one of the following instruments is to be mounted on or under a 940 Professional IC Vario:

- One or more 942 Extension Module Vario
- Or another instrument with the same-sized footprint

4.2.2 Mounting base tray and bottle holder (optional)

The base tray and bottle holder come fully assembled on a new ion chromatograph. To install an Extension Module on the ion chromatograph, remove the bottle holder and put it back on top of the topmost instrument. To install an Extension Module under the ion chromatograph, remove the base tray and set it under the lowest instrument.

4.2.2.1 Removing/mounting the base tray

Remove the base tray to install another instrument under the IC instrument.



CAUTION

Do not pinch capillaries or leak sensor cables

Capillaries are fed through the guide ducts between the base tray and the instrument. Pinches in the leak sensor cable or the capillaries may lead to malfunctions.

- Unplug the leak sensor cable before you remove the base tray.
- Remove all the capillaries from the capillary ducts before you remove the base tray.

Removing the base tray

Prerequisites

- The instrument is switched off.
- The bottle holder is cleared.
- All of the cable connections on the rear have been disconnected.
- The capillaries are removed from the guide ducts between the instrument and the base tray.
- There are no loose parts in the instrument.

Accessories

- 3 mm hex key (6.2621.100)




Figure 3 Removing the base tray

- 1** Tilt the instrument sideways and lay it down flat.
- 2** Loosen the 4 cylinder screws with the 3 mm hex key. Remove the cylinder screws and their washers.
- 3** Remove the base tray.

Always mount the base tray under the lowermost instrument of the stack.

Mounting the base tray

Prerequisites

- The instrument is switched off.
- The bottle holder is cleared.
- All of the cable connections on the rear have been disconnected.
- There are no loose parts in the instrument.
- The instrument is lying on its side, and the bottom surface is visible.

Accessories

- 3 mm hex key (6.2621.100)




Figure 4 Mounting the base tray

- 1 Mount the base tray in such a way that the openings in the base tray match exactly the screw threads in the bottom of the instrument.
- 2 Slide the washers onto the cylinder screws. Insert the cylinder screws with the washers and tighten them with the 3 mm hex key.
- 3 Set the instrument back up on the base tray.

Stack other instruments in the required order. Mount the bottle holder (6.2061.100) onto the topmost instrument on the stack (see "Mounting the bottle holder", page 14).

4.2.2.2 Removing/mounting the bottle holder

Remove the bottle holder if you want to install another instrument onto the IC instrument.

Removing the bottle holder

Prerequisites

- The instrument is switched off.
- The bottle holder is cleared.
- Drainage tubing is disconnected from the drainage tubing connection of the bottle holder.
- The capillaries are removed from the guide ducts between the instrument and the bottle holder.

Accessories

- 3 mm hex key (6.2621.100)




Figure 5 Removing the bottle holder

- 1 Remove the 2 covering stoppers.
- 2 Loosen the 2 cylinder screws with the 3 mm hex key and remove them.
- 3 Remove the bottle holder.

Stack other instruments in the required order. Mount the bottle holder (6.2061.100) onto the topmost instrument on the stack.

Mounting the bottle holder

Prerequisite

- The instrument is switched off.

Accessories

- 3 mm hex key (6.2621.100)




Figure 6 Mounting the bottle holder

- 1 Mount the bottle holder onto the topmost instrument in such a way that the openings in the bottle holder exactly match the screw threads on the top surface of the instrument.
- 2 Insert the 2 cylinder screws and tighten them with the 3 mm hex key.
- 3 Insert both covering stoppers.

After attaching the bottle holder, restore all connections that were loosened at the beginning of the process.

Restoring the loosened connections

- 1 Plug in all necessary USB cables.
- 2 Plug in all necessary MSB cables.
- 3 Plug in the power cord.
- 4 Mount the drainage tubing again (*see manual of the IC instrument*).
A longer section of silicone tubing (6.1816.020) may have to be cut to size and mounted (*see also the manual for the IC instrument*).
- 5 If one of the instruments in the stack is equipped with a leak sensor connection socket, connect the leak sensor (*see manual of the IC instrument*).
- 6 Restore any capillary connections that may have been removed.



4.3 Connecting the eluent bottle

For the automated eluent production, the eluent bottle must be equipped with a bottle cap for a level sensor and a sensor of the type "FULL".

Inserting the measuring sensor

Accessories

For this step you need:

- A bottle cap for level sensor (6.1626.000)
- FULL sensor (6.2769.000) for anion eluents or EMPTY sensor (6.2769.1X0) for cation eluents




Figure 7 Bottle cap for level sensor - Inserting the measuring sensor

1 Measuring sensor opening

For two measuring sensors (inside) and the level sensor cable (outside).

2 M6 opening

For the FEP tubing connection to the Dosino.

3 Adsorber opening

4 M8 opening

For the capillary connection to the ion chromatograph.

- 1 Remove the stoppers from the bottle cap.
- 2 Place the bottle cap on a surface with the opening facing up.
- 3 Insert the two measuring sensors from the inside into the two openings for measuring sensors.

Assembling the bottle cap

Accessories

For this step you need:

- Eluent bottle (6.1608.070) from the accessories set of the ion chromatograph.

- Bottle cap for level sensor (6.1626.000) with inserted measuring sensors (see "Inserting the measuring sensor", page 16).
- FEP aspiration tubing (6.1819.100)
- FEP tubing (6.1805.530)
- Level sensor cable (6.2151.060)
- Adsorber tube (6.1619.000)
- Spiral band (6.1815.010)




Figure 8 Installing the eluent bottle cap with level sensor

1 Connecting the FEP tubing

Fasten the tubing connection to the Dosino to the M6 opening.

- Insert the FEP aspiration tubing into the M6 opening.
- Screw the FEP tubing into the M6 opening.

2 Connecting the eluent aspiration tubing

Fasten the capillary connection to the ion chromatograph to the M8 opening.

- Start by pushing the M8 tubing nipple over the end of the aspiration tubing, followed by the O-ring.
- Temporarily secure the eluent aspiration tubing in the M8 opening.



- Attach the tubing weighting and the aspiration filter to the end of the eluent aspiration tubing, see the manual of the ion chromatograph.

3 Connecting the level sensor cable

Connect the two plugs of the level sensor cable to the two sockets.

4 Mounting the adsorber

Fill the adsorber tube with adsorber material and cotton and insert it into the last remaining opening.

5 Bundling the level sensor cables and tubings



NOTICE

The fine connection cables of the level sensor cable break off easily if the bottle cap is rotated too much.

The level sensor cable is damaged as a result and needs to be replaced.

To prevent the connection cables from breaking off, we recommend bundling the level sensor cable together with the tubings and capillaries that are connected to the bottle cap.

Tie the level sensor cable, the FEP tubing and the eluent aspiration tubing together using a spiral band. Tie a piece of spiral band with a length of approx. 10 cm around the cable and the tubings as close as possible to the bottle cap.

6 Attaching the bottle cap

Screw the bottle cap to the eluent bottle and install the eluent bottle on the bottle holder of the ion chromatograph.

4.4 Connecting the concentrate bottle

The bottle with the concentrate can be equipped with a bottle cap without level sensor, e.g. with the bottle cap (6.1602.160).

Accessories

For this step you need the following accessories:

- Bottle (6.1608.070) filled with concentrate
- Bottle cap (6.1602.160)
- FEP aspiration tubing (6.1819.100)
- FEP tubing (6.1805.530)
- Adsorber tube (6.1609.000)
- SGJ clip for SGJ 14/15 (6.2023.020)




Figure 9 Installing the concentrate bottle cap without level sensor

1 Connecting the FEP tubing

Fasten the tubing connection to the Dosino to the M6 opening.

- Insert the FEP aspiration tubing into the M6 opening.
- Screw the FEP tubing into the M6 opening.

2 Sealing the M8 opening

Screw the M8 threaded stopper into the M8 opening.

3 Mounting the adsorber

Fill the adsorber tube with adsorber material and cotton (for anion eluents) or only with cotton (for cation eluents) and insert it into the large opening. Secure in place with an SGJ clip.



Recommendation

If you intend to let your ion chromatography system run unsupervised for extended periods of time, we recommend also monitoring the level of the concentrate bottle with a level sensor.

For this purpose, you need to order the following optional accessories:

- Bottle cap for level sensor (6.1626.000) incl. adsorber tube (6.1619.000)
- EMPTY sensor (6.2769.110 for 2 L bottle or 6.2769.100 for 1 L bottle or 6.2769.120 for 10 L canister)
- Level sensor cable (6.2151.060)
- FEP tubing (6.1805.530), is contained in the accessories of the 941 Eluent Production Module.




Figure 10 Installing the concentrate bottle cap with level sensor

1 Inserting the measuring sensor

Insert the two measuring sensors from the inside into the two openings for measuring sensors.

2 Connecting the FEP tubing

Fasten the tubing connection to the Dosino to the M6 opening.

- Insert the FEP aspiration tubing (6.1819.100) into the M6 opening.
- Screw the FEP tubing into the M6 opening.

3 Sealing the M8 opening

Screw the M8 threaded stopper into the M8 opening.

4 Connecting the level sensor cable

Connect the two plugs of the level sensor cable to the two sockets.

5 Mounting the adsorber

Fill the adsorber tube with cotton and insert it into the last remaining opening.

6 Bundling the level sensor cables and tubings**NOTICE**

The fine connection cables of the level sensor cable break off easily if the bottle cap is rotated too much.

The level sensor cable is damaged as a result and needs to be replaced.

To prevent the connection cables from breaking off, we recommend bundling the level sensor cable together with the tubings and capillaries that are connected to the bottle cap.

Tie the level sensor cable and the FEP tubing together using a spiral band. Tie a piece of spiral band with a length of approx. 10 cm around the cable and the tubings as close as possible to the bottle cap.

7 Attaching the bottle cap

Screw the bottle cap onto the concentrate bottle.



4.5 Connecting the ultrapure water container

The ultrapure water can be aspirated either from a 10 L canister or an ELGA PURELAB water purification system.

The 10 L canister is included in the scope of delivery and is used for the standard installation.

4.5.1 Connecting the ultrapure water canister

Assembling the bottle cap without level sensor

The canister with the ultrapure water can be equipped with a bottle cap without level sensor, e.g. with the bottle cap (6.1602.160).

Accessories

For this step you need the following accessories:

- Canister, 10 L (6.1621.000)
- Bottle cap (6.1602.160)
- FEP aspiration tubing (6.1819.100)
- FEP tubing (6.1805.530)
- Adsorber tube (6.1609.000 + 6.1624.000)




Figure 11 Installing the ultrapure water canister cap without level sensor

1 Connecting the FEP tubing

Fasten the tubing connection to the Dosino to the M6 opening.

- Insert the FEP aspiration tubing into the M6 opening.
- Screw the FEP tubing into the M6 opening.

2 Sealing the M8 opening

Screw the M8 threaded stopper into the M8 opening.

3 Mounting the adsorber

Fill the adsorber tube with soda lime and cotton and insert it into the large opening. Secure in place with an SGJ clip.

Option: Assembling the bottle cap with level sensor

If you intend to let your ion chromatography system run unsupervised for extended periods of time, we recommend also monitoring the ultrapure water canister with a level sensor.

Accessories

For this step you need the following optional accessories:

- A bottle cap for level sensor (6.1626.000)
- EMPTY sensor for 10 L canister (6.2769.120)




Figure 12 Installing the ultrapure water canister cap with level sensor

1 Inserting the measuring sensor

Insert the two measuring sensors from the inside into the two openings for measuring sensors.



2 Connecting the FEP tubing

Fasten the tubing connection to the Dosino to the M6 opening.

- Insert the FEP aspiration tubing (6.1819.110) into the M6 opening.
- Screw the FEP tubing into the M6 opening.

3 Sealing the M8 opening

Seal the M8 opening with a stopper.

4 Connecting the level sensor cable

Connect the two plugs of the level sensor cable to the two sockets.

5 Mounting the adsorber

Fill the adsorber tube with soda lime and cotton and insert it with the adapter into the last remaining opening.

6 Bundling the level sensor cables and tubings



NOTICE

The fine connection cables of the level sensor cable break off easily if the bottle cap is rotated too much.

The level sensor cable is damaged as a result and needs to be replaced.

To prevent the connection cables from breaking off, we recommend bundling the level sensor cable together with the tubings and capillaries that are connected to the bottle cap.

Tie the level sensor cable and the FEP tubing together using a spiral band. Tie a piece of spiral band with a length of approx. 10 cm around the cable and the tubings as close as possible to the bottle cap.

7 Attaching the bottle cap

Screw the bottle cap onto the canister.

4.5.2 Connecting an ELGA PURELAB flex

The 941 Eluent Production Module can also be connected directly to an ELGA PURELAB flex water purification system, as an alternative to the ultrapure water canister.

The ELGA PURELAB flex is not included in the scope of delivery and has to be ordered separately.

An ultrapure water insert specifically for Metrohm applications is included in the ELGA PURELAB flex. This insert ensures that ultrapure water is aspirated at all times and has two connectors, each for one piece of FEP tubing.

This ultrapure water insert must be put in place during installation.

Accessories

For this step you need the following accessories:

- ELGA PURELAB flex, incl. ultrapure water insert
- FEP tubing (6.1805.530)




Figure 13 Installing an ELGA PURELAB flex

1 Composite aeration filter

2 Ultrapure water insert

Installing an ELGA PURELAB flex

Install the ELGA PURELAB flex as described in the manufacturer's documentation.



Carry out the following steps before installing the composite aeration filter.

1 Installing the ultrapure water insert

Screw the ultrapure water insert supplied in the opening in which the composite aeration filter is to be installed according to the manufacturer's documentation.

2 Inserting the composite aeration filter

Screw the composite aeration filter onto the ultrapure water insert.

3 Connecting the FEP tubing

Screw the FEP tubing onto one of the connectors of the ultrapure water insert.

4.6 Installing the Dosino

Attaching the Dosino to the Dosing Unit

Accessories

For this step you need the following accessories:

- 800 Dosino (2.800.0010)
- Dosing Unit 50 mL (6.3032.250)

The manual for the 800 Dosino describes how to attach the 800 Dosino correctly to a dosing unit.



CAUTION

Please read through the correct procedure in the manual for the 800 Dosino before you attach the Dosino to the dosing unit.

Installing FEP tubings

Accessories

For this step you need the following accessories:

- 800 Dosino (2.800.0010) with dosing unit 50 mL (6.3032.250) attached
- The three FEP tubings connected to the bottles, one additional FEP tubing (6.1805.530)
- Adapter Dosino port 4 (6.1808.280)




Figure 14 Connecting the Dosino

1 Port 1

For connecting the concentrate bottle.

2 Port 2

For connecting the ultrapure water supply.

3 Port 3

For connecting the eluent bottle.

4 Port 4

For connecting the waste container.

1 Connecting the concentrate bottle

- Push the FEP tubing attached to the concentrate bottle through the cable conduit from the rear to the front of the 941 Eluent Production Module.
- Screw the FEP tubing onto Port 1 of the Dosino.

2 Connecting the ultrapure water

- Push the FEP tubing connected to the ultrapure water canister or the ELGA PURELAB flex through the cable conduit from the rear to the front of the 941 Eluent Production Module.
- Screw the FEP tubing onto port 2 of the Dosino.



3 Connecting the eluent bottle

- Push the FEP tubing attached to the eluent bottle through the cable conduit from the rear to the front of the 941 Eluent Production Module.
- Screw the FEP tubing onto port 3 of the Dosino.

4 Connecting the waste container

- Attach the adapter to port 4 of the Dosino and screw the fourth FEP tubing onto the adapter.
- Guide the other end of the FEP tubing from the front through the cable conduit to the rear of the 941 Eluent Production Module and attach to the waste container.


Inserting the Dosino into the instrument

- 1 Insert the Dosino sideways into the Dosino holder.

The cable has to rest on the left side in the recess designated for this purpose.

- 2 Guide the Dosino cable through the cable conduit to the rear of the instrument.

- 3 Plug the Dosino cable into one of the MSB sockets.



4.7 Connecting the instrument

4.7.1 Connecting the instrument to the ion chromatograph



NOTICE

The 940 Professional IC Vario must be switched off when connecting the 941 Eluent Production Module!

Accessories

For this step, you need the following accessories:

- Controller cable (6.2151.000)

- 1 Insert the Mini DIN plug of the controller cable into the connection socket on the rear of the instrument labeled *Controller*.
- 2 Insert the USB A plug of the controller cable into a USB socket of the ion chromatograph.

4.7.2 Connecting the instrument to the power grid



WARNING

Electric shock from electrical potential

Risk of injury by touching live components or through moisture on live parts.

- Never open the housing of the instrument while the power cord is still connected.
- Protect live parts (e.g. power supply unit, power cord, connection sockets) against moisture.
- Unplug the power plug immediately if you suspect that moisture has gotten inside the instrument.
- Only personnel who have been issued Metrohm qualifications may perform service and repair work on electrical and electronic parts.

Connecting the power cord

Accessories

Power cord with the following specifications:

- Length: max. 2 m
- Number of cores: 3, with protective conductor
- Instrument plug: IEC 60320 type C13
- Conductor cross-section 3x min. 0.75 mm² / 18 AWG



- Power plug:
 - according to customer requirement (6.2122.XX0)
 - min. 10 A



NOTICE

Do not use a not permitted power cord!

1 Plugging in the power cord

- Plug the power cord into the instrument's power socket.
- Connect the power cord to the power grid.

5 Operation

The instrument is operated via MagIC Net software only. Additional information on operating MagIC Net can be found in the document "*MagIC Net Tutorial*" or in the software's online help.

6.3 Dosino

Information regarding operation and maintenance of the Dosino can be found in the manual for the 800 Dosino.



NOTICE

Please read the chapter "Operation and maintenance" in the manual for the 800 Dosino and follow the instructions described in it.

6.4 Dosing unit

Information regarding operation and maintenance of the dosing unit can be found in the manual for the 807 Dosing Unit.



NOTICE

Please read the chapter "Operation and maintenance" in the manual for the 807 Dosing Unit and follow the instructions described in it.

8 Technical specifications

8.1 Reference conditions

The technical specifications listed in this chapter refer to the following reference conditions:

<i>Ambient temperature</i>	+25 °C (± 3 °C)
<i>Relative humidity</i>	$\leq 60\%$
<i>Instrument status</i>	> 30 minutes in operation
<i>Adjusting interval</i>	annual

8.2 Ambient temperature

Automatic interior temperature monitoring

<i>Nominal function range</i>	+5 - +45 °C at a maximum of 85% humidity
<i>Storage</i>	-20 - +60 °C
<i>Transport</i>	-40 - +60 °C

8.3 Level sensor

Measuring inputs

<i>Switching threshold</i>	1 μ S
<i>Measurement frequency</i>	Approx. 1.7 kHz



8.4 Housing

Dimensions

Width 365 mm

Height 131 mm

Depth 380 mm

Material of base tray, housing and bottle holder Polyurethane hard foam (PUR) with flame retardation for fire class UL94V0, CFC-free, coated

8.5 Weight

2.941.0010 7.51 kg

8.6 Dosino

You can find the technical specifications for the Dosino in the manual for the 800 Dosino.

8.7 Dosing unit

You can find the technical specifications for the dosing unit in the manual for the 807 Dosing Unit.

8.8 Power connection

Required supply voltage 100 - 240 V \pm 10% (autosensing)

Required frequency 50 - 60 Hz \pm 3%

Power consumption 45 W maximum

Fuse Electronic overload protection

8.9 Interfaces

Controller

Input

1 8-pin Mini DIN socket for a connecting cable to a computer or an ion chromatograph for controlling the instrument.

USB

Number

2 type A USB sockets for connecting Metrohm instruments or USB peripheral devices of other manufacturers.

MSB

Number

3 Mini DIN sockets (8-pin) for connecting Dosinos (MSB 1 is used internally).

Level sensor

Measuring input

4 Mini DIN sockets (8-pin) for connecting level sensor cables.

Index

A

Ambient temperature 35

B

Base tray

Mount 12

Remove 11

Bottle holder

Mount 14

Remove 13

C

Care 32

Concentrate bottle

Connect 19

Connect

Concentrate bottle 19

ELGA PURELAB flex 25

Eluent bottle 16

Power grid 29

To IC instrument 29

To power grid 36

Ultrapure water canister 22

Water purification system ... 25

Connector (Controller) 29

Controller 37

D

Dimensions 36

Door 32

Dosino

Install 26

E

Electrostatic charge 4

Eluent bottle

Connect 16

F

Frequency 36

Fuse 36

H

Housing 36

I

Instrument

Connect 29

Interface

Controller 37

MSB 37

USB 37

Interfaces 37

L

Level sensor

Install 16

M

Material 36

MSB 37

P

Power connection 29, 30, 36

Power consumption 36

R

Reference conditions 35

S

Safety instructions 3

Service 3

Storage 35

Supply voltage 3, 36

T

Technical specifications

Interfaces 37

Reference conditions 35

Transport 35

Troubleshooting

Dosing unit 34

Dosino 34

U

Ultrapure water canister

Connect 22

USB 37

W

Water purification system

Connect 25