

Application Bulletin

Of interest to:

Environmental protection; Electroplating

A 1, 2, 10

Potentiometric determination of cyanide

Summary

The determination of cyanide is very important not only in electroplating baths and when decontaminating waste water but, due to its high toxicity, also in water samples in general. Concentrations of 0.05 mg/L CN⁻ can already be lethal for fish.

This bulletin describes the determination of cyanide in samples of different concentrations by means of potentiometric titration.

Chemical reactions:

$$2 \text{ CN}^- + \text{Ag}^+ \rightarrow [\text{Ag}(\text{CN})_2]^-$$

 $[\text{Ag}(\text{CN})_2]^- + \text{Ag}^+ \rightarrow 2 \text{ AgCN}$

Instruments and accessories

- 702 SET/MET Titrino, 716 DMS Titrino, 736 GP Titrino, 751 GPD Titrino or 785 DMP Titrino or 726 Titroprocessor with 700 Dosino or 685 Dosimat
- 2.728.0040 Magnetic Stirrer
- 6.3014.223 and/or 6.3014.213 Exchange Units
- 6.0430.100 Ag Titrode with Ag₂S coating and 6.2104.020 electrode cable

Reagents

- Titrant c(AgNO₃) = 0.1 mol/L ... 0.0002 mol/L
- Sodium hydroxide c(NaOH) = 1 mol/L
- Digestion chemicals: see under «3. Cyanide traces in water samples after sample preparation»

Analysis

1. Electroplating baths (high cyanide concentrations): alkaline baths for Ag, Cd, Cu, Pb, Zn, etc.

Pour ca. 50 mL dist. water as well as 5 mL c(NaOH) = 1 mol/L into a glass beaker. Add 2.0 mL of the plating bath sample and titrate with $c(AgNO_3) = 0.1 \text{ mol/L}$ just past the first equivalence point.



Calculation

1 mL c(AgNO $_3$) = 0.1 mol/L corresponds to 2.6018 mg CN $^-$ or 6.5116 mg KCN or 4.9007 mg NaCN

g/L cyanide = EP1 * C01 / C00

EP1 = titrant consumption to reach the first EP in mL

C00 = 2.0 (sample volume in mL)

C01 = 2.6018 or 6.5116 or 4.9007 (equivalent weight of CN⁻ or KCN or NaCN in mg/mL)

2. Waste water for decontamination (1 ... 100 mg/L CN)

Pour 5 mL c(NaOH) = 1 mol/L into a glass beaker. Add 10 ... 100 mL sample solution (depending on the expected CN^- concentration) and titrate with c(AgNO₃) = 0.01 mol/L just past the first equivalence point.

Calculation

1 mL c(AgNO₃) = 0.01 mol/L corresponds to 0.2602 mg CN⁻

 $mg/L CN^{-} = EP1 * C01 * C02 / C00$

EP1 = titrant consumption to reach the first EP in mL

C00 = 10 ... 100 (sample volume in mL)

C02 = 0.2602 (equivalent weight of CN^- in mg/mL)

C03 = 1000 (conversion factor in mL/L)

3. Cyanide traces in water samples after sample preparation

The decomposition and separation of the cyanides are carried out according to DIN 38405 D13/D14).

One differentiates between the easily released cyanides and the total cyanide content.

Apparatus for decomposition and separation of the cyanides

The apparatus consists of a washing bottle, a three-necked round flask, a condenser and an absorption vessel (see Fig. 1).

3.1 Determination of the easily released cyanides

These are released from the sample as HCN at room temperature and a pH value of about 4 by means of an air stream and absorbed in sodium hydroxide solution.

Reagents

- Sodium hydroxide c(NaOH) = 1 mol/L (for the absorption vessel and washing bottle)
- Hydrochloric acid c(HCI) = 1 mol/L
- ZnSO₄/CdSO₄ solution:
 Dissolve 100 g ZnSO₄ * 7 H₂O as well as 100 g CdSO₄ * 8 H₂O in dist. water and make up to 1 L.



- EDTA solution:
 Dissolve 100 g Na₂EDTA * 2 H₂O in 940 mL dist. water with heating.
- Buffer solution pH = 4.0:
 Dissolve 80 g potassium hydrogen phthalate in 920 mL dist. water with heating.
- Zinc powder, p.a.

Separation of the easily released cyanides

Fill the absorption vessel with 10 mL c(NaOH) = 1 mol/L, mount it directly onto the three-necked flask and connect it to the suction tubing. Add through the charging funnel 10 mL EDTA solution, 10 mL ZnSO $_4$ /CdSO $_4$ solution, 50 mL buffer solution pH = 4 as well as 100 mL water sample, then mix well. If necessary, the pH value of the mixture is adjusted to 3.9 ± 0.1 by adding c(HCI) = 1 mol/L or c(NaOH) = 1 mol/L drop by drop through the funnel. Afterwards add 0.3 zinc powder through the side neck opening, then close this again. Connect the washing bottle containing ca. 100 mL c(NaOH) = 1 mol/L and switch on the suction pump. Set the air flow at 60 L/h and suck air through the apparatus for 4 h.

With water samples containing less than $0.1~\text{mg/L}~\text{CN}^-$ a sample volume of 200 mL is used. In this case the additions of EDTA, ZnSO₄/CdSO₄ and buffer solution have to be doubled, too.

It is necessary to determine the blank of the chemicals used. Instead of the sample, take 100 mL dist. water for this analysis. The blank determination has to be repeated whenever new packages of chemicals are opened for use.

Analysis

The contents of the absorption vessel are rinsed into a glass beaker with ca. 20 mL dist. water.

For CN $^-$ concentrations of 0.2 ... 10 mg/L, titrate with c(AgNO $_3$) = 0.002 mol/L, for CN $^-$ concentrations of 0.01 ... 0.5 mg/L titrate with c(AgNO $_3$) = 0.0002 mol/L using the following parameter settings:

 $\begin{array}{lll} \text{meas.pt.density} & 6 \\ \text{min.incr.} & 10.0 \ \mu\text{L} \\ \text{titr.rate} & \text{max.} \\ \text{signal drift} & 20.0 \ \text{mV/min} \\ \text{equilibr.time} & 20 \ \text{s} \\ \end{array}$

equilibr.time 20 s pause 60 s EPC 20

In order to condition the electrode, it is immersed before the first analysis for 20 min in sodium hydroxide solution with a concentration of c(NaOH) = 0.4 mol/L that additionally contains 0.3 mg/L CN^- .

Calculation

1 mL c(AgNO₃) = 0.002 mol/L corresponds to 0.052 mg CN $^-$ 1 mL c(AgNO₃) = 0.0002 mol/L corresponds to 0.0052 mg CN $^-$

 $mg/L CN^{-} = (EP1 - C31) * C01 * C02 * C30 / C00$

EP1 = titrant consumption in mL

C00 = 100 ... 200 (sample volume in mL)

C01 = 0.052 or 0.0052 (equivalent weight of CN⁻ in mg/mL)

C02 = 1000 (conversion factor in mL/L)



C30 = titer of the titrant

C31 = titrant consumption for the blank in mL

Remarks

The titer of the titrant is determined with 0.1 mg [when using c(AgNO₃) = 0.002 mol/L] or 0.05 mg CN⁻ standard [when using c(AgNO₃) = 0.0002 mol/L] in 30 mL c(NaOH) = 0.4 mol/L:

titer = 1.923 / EP1 or titer = 9.615 / EP1

Low CN⁻ concentrations lead to a slow initial response of the electrode. Therefore a pause of 60 s is programmed on the titrator.

3.2 Determination of the total cyanide content

The cyanide compounds are decomposed by boiling with HCI in the presence of Cu(I) ions. The HCN formed is expelled with an air stream and absorbed in sodium hydroxide solution.

Reagents

- Sodium hydroxide c(NaOH) = 1 mol/L (for the absorption vessel and washing bottle)
- Concentrated hydrochloric acid w(HCI) = 35 ... 37%
- CuSO₄ solution:
 Dissolve 200 g CuSO₄ * 5 H₂O in dist. water and make up to 1 L.
- Sn(II) solution:
 Dissolve 50 g SnCl₂ * 2 H₂O as well as 40 mL c(HCl) = 1 mol/L in dist. water and make up to 100 mL. The solution is stable for about one week.

Decomposition and separation of the total cyanide

The absorption vessel is filled with 10 mL c(NaOH) = 1 mol/L, mounted onto the reflux cooler and connected to the suction tubing. Through the charging funnel add the following components one after the other: ca. 30 mL dist water, 10 mL $CuSO_4$ solution, 2 mL Sn(II) solution and finally 100 mL water sample. After addition of 10 mL conc. HCI, connect the washing bottle containing ca. 100 mL c(NaOH) = 1 mol/L to the funnel and heat up the contents of the flask to boiling point. Set the air flow at ca. 20 L/h; the reflux should be ca. 1 ... 2 drops/s. The decomposition is completed after 1 h.

With water samples containing less than 0.1 mg/L CN⁻ a sample volume of 200 mL is used. In this case the additions of CuSO₄ solution, Sn(II) solution and conc. HCl have to be doubled, too.

It is necessary to determine the blank of the chemicals used. Instead of the sample, take 100 mL dist. water for this analysis. The blank determination has to be repeated whenever new packages of chemicals are opened for use.

Analysis

The contents of the absorption vessel are rinsed into a glass beaker with ca. 20 mL dist. water. The analysis is then carried out in the same way as described under «3.1 Determination of the easily released cyanides».



Figures





Fig. 1: Apparatus for decomposition and separation of the cyanides.



'pa		
785 DMP Titrino	02287	785.0010
date 1999-07-16	time 11:0	07 5
DET U	*****	
parameters		
>titration paramete	ers	
meas.pt.density	4	
min.incr.	10.0	
dos.rate	max.	ml/min
signal drift	50	mV/min
equilibr.time	26	s
start V:	OFF	
pause	0	S
meas.input:	1	
temperature	25.0	°C
>stop conditions		
stop V:	abs.	
stop V	3	ml
stop U	OFF	mV
stop EP	9	
filling rate	max.	ml/min
>statistics		
status:	OFF	
>evaluation		
EPC	30	
EP recognition:	all	
fix EP1 at U	OFF	mV
pK/HNP:	OFF	
>preselections		
req.ident:	OFF	
req.smpl size:	OFF	
limit smpl size:	OFF	
activate pulse:	OFF	

Fig. 2: Parameter settings on the 785 DMP Titrino for the determination of cyanide in waste water before decontamination.






Fig. 3: Result block and titration curve.



'pa 785 DMP Titrino date 1999-07-19 DET U		785.0010 03 9
parameters		
>titration paramete	ers	
meas.pt.density	6	
min.incr.	10.0	ul
dos.rate		ml/min
signal drift		mV/min
equilibr.time	20	•
start V:	OFF	
pause	60	s
meas.input:	1	
temperature	25.0	°C
>stop conditions		
stop V:	abs.	
stop V	3	ml
stop U	OFF	mV
stop EP	9	
filling rate	max.	ml/min
>statistics		
status:	OFF	
>evaluation		
EPC	73	
EP recognition:	all	
fix EP1 at U	OFF	mV
pK/HNP:	OFF	
>preselections		
req.ident:	OFF	
req.smpl size:	OFF	
limit smpl size:	OFF	
activate pulse:	OFF	

'fr 785 DMP Titrino date 1999-07-19 card label:Appl.751		
U(init) 48 smpl size 100	ml	*****
EP1 1.208 CN- 0.63 stop V reached		-225 mV
scop v reached		
'cu		
785 DMP Titrino date 1999-07-19		
start V 0.000 r 2.0 ml/div dU=1		*****




Fig. 4: Parameter settings on the 785 DMP
Titrino for the determination of cyanide
traces in surface water after decomposition according to DIN 38405 part 14.

Fig. 5: Result block and titration curve.

Literature

- Metrohm Application Note T-22
 Cyanide in alkaline plating baths for cadmium, copper, lead or zinc Metrohm Ltd., Herisau.
- DIN 38405, part 13 Anionen (Gruppe D). Bestimmung von Cyaniden.
- DIN 38405, part 14
 Anionen (Gruppe D). Bestimmung von Cyaniden in Trinkwasser, gering belastetem Grund- und Oberflächenwasser.