

Determination of Anions in Municipal Drinking Water Using a Compact Ion Chromatography System

Terri Christison and Jeff Rohrer, Thermo Fisher Scientific, Sunnyvale, CA, USA

Key Words

HPIC, Integriion, Dionex IonPac AS22-Fast-4 μ m Column, Dionex AS22 Eluent Concentrate, Dionex AERS 500 Suppressor, Inorganic Anions, Water Analysis

Introduction

This application proof note demonstrates a 5-minute, fast separation of anions in a municipal drinking water sample on the 4 μ m resin particle format Thermo Scientific™ Dionex™ IonPac™ AS22-Fast-4 μ m column. This method upgrades the method published in [Application Brief 120](#)¹ by improving the separation using the 4 μ m resin particle column to produce more efficient peaks. In this proof note, the method is performed using a Thermo Scientific™ Dionex™ Integriion™ HPIC™ system, which works with the high pressures generated by the 4 μ m particles to deliver improved separation.

Method

IC System:	Thermo Scientific Dionex Integriion HPIC system
Columns:	Thermo Scientific™ Dionex™ IonPac™ AS22-Fast-4 μ m Analytical (2 × 150 mm) Thermo Scientific Dionex IonPac AG22-Fast-4 μ m Guard (2 × 30 mm)
Eluent:	4.5 mM Sodium Carbonate/1.4 mM Bicarbonate
Flow Rate:	0.5 mL/min
Injection Volume:	2.5 μ L
Temperature:	30 °C
Detection:	Suppressed conductivity, Thermo Scientific™ Dionex™ AERS™ 500 Carbonate Suppressor, 2 mm, 17 mA, recycle mode

Reference

1. Thermo Scientific Application Brief 120: Municipal Drinking Water Analysis by Fast IC, Sunnyvale, CA [Online] <http://www.thermoscientific.com/content/dam/tfs/ATG/CMD/CMD%20Documents/Application%20&%20Technical%20Notes/Chromatography/GC%20HPLC%20and%20UHPLC%20Columns%20and%20Accessories/Chromatography%20Column%20Accessories/109811-AB120-IC-Drinking-Water-CarbonateEluents-19Jan2011-LPN2712.pdf> (accessed Jan. 14, 2016)

For application support, visit the [AppsLab Library](#) where you can find detailed method information, chromatograms and related compound information. All the information needed to run, process and report the analysis is available in ready-to-use eWorkflows, which can be executed directly in your chromatography data system. www.thermoscientific.com/appsab




Figure 1. Determination of anions in a municipal drinking water sample.

www.thermoscientific.com/integriion

©2016 Thermo Fisher Scientific Inc. All rights reserved. ISO is a trademark of the International Standards Organization. All other trademarks are the property of Thermo Fisher Scientific and its subsidiaries. This information is presented as an example of the capabilities of Thermo Fisher Scientific products. It is not intended to encourage use of these products in any manners that might infringe the intellectual property rights of others. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

AB71908-EN 0116S



Thermo Fisher Scientific,
Sunnyvale, CA USA is
ISO 9001 Certified.

Thermo
SCIENTIFIC

A Thermo Fisher Scientific Brand