



The Acclaim PolarAdvantage II column has better selectivity for these compounds than C18; the unknown and peak 2 coelute on C18. The RSLC column is three times faster and uses only 16% of the solvent, yet produces the same quality of chromatogram.