Starting primary compound and its purity:



Analytika, spol. s r.o., Oddělení RM, výrobce RM č. 7501, akreditovaný ČIA podle ČSN EN ISO 17034:2017 Analytika, spol. s r.o., Department of RM, Reference Materials Producer No. 7501 accredited by CAI according to ČSN EN ISO 17034:2017

# CERTIFICATE

### **Aqueous calibration solution**

# ASTASOL<sup>®</sup> AN90731HIC

This Certificate is designed in accordance with ISO Guide 31

Category:

#### Analyte:

## Fluorides (F<sup>-</sup>)

**Certified reference material** 

Product code:

NaF, 99.99%

**AN90731HIC** 

Matrix:

Ultrapure demineralized water (resistivity  $\geq 18M\Omega.cm$ , 0.22µm filtered)

The solution is filtered through a membrane ultrafilter 0.45  $\mu$ m to avoid clogging of a chromatography column.

**Density and its expanded uncertainty** ( $\mathbf{k} = 2$ ): 1.0012 ± 0.0005 g/cm<sup>3</sup> (at 20 °C)

Certified value of concentration and its expanded uncertainty (k = 2) at 20  $^{\circ}C$ 

### $1 \ 000 \pm 2 \ mg/l$

 $999 \pm 3 \text{ mg/kg}^*$ 

\*Mass fraction in mg/kg is derived from density

**Specification:** 

Batch No.: 9009

The date of production: 09.06.2023

Shelf life: 5 years from the date of production

The date of first opening of the aluminium bag: .....

Verze certifikátu / Version of the certificate: 01

1/3

Verze formuláře / Version of the form: 03

ANALYTIKA®, spol. s r.o., Oddělení referenčních materiálů / Department of reference materials Ke Klíčovu 2a/816, CZ-190 00 Praha 9 – Vysočany, Tel/Fax: +420 286 589 616, sales@analytika.net, www.analytika.net

ه) analytika

#### Intended use:

For calibration and validation of analytical methods analysing aqueous solutions such as atomic spectrometry (AAS, AES, ICP-OES, ICP-MS, IC), molecular absorption spectrometry and selected electroanalytical methods

#### **Certification and traceability:**

This CRM is certified on the basis of gravimetric preparation. This procedure also ensures a direct traceability to SI unit - kg. Certified concentration, its uncertainty and traceability were verified by complexometric titration (SRM NIST 928).

#### Trace impurities in bottled solution (in mg/l):

Determination of trace impurities was performed by ion chromatography. Impurity levels are supplied only for information of the user and should not be used as calibration data.

Bromides	Chlorides	Fluorides	Iodides	Nitrites	Nitrates	Phosphates	Sulfates
<0,02	<0,02	А	<0,02	<0,02	<0,02	<0,02	<0,02
< v - ballow d	lataction limit	$\Lambda - on obtained$	uto				

< x = bellow detection limit A = analyte

#### Homogeneity and stability:

It has been demonstrated that this CRM is homogeneous and its stability is guaranteed during the whole shelf life provided the solution it kept under conditions presented below.

#### Storing and instruction for use:

This CRM must be stored in the original closed bottle between 5 - 30 °C. The producer guarantees a declared shelf life and expiration time provided the CRM is properly stored and professionally handled. The temperature of the solution must be  $20 \pm 0.5$  °C before every use. It is necessary to indicate on this certificate and the label the expiration time, which depends on the date of the first time the aluminium bag was opened. After use, the bottle must be immediately tightly capped, and it is recommended to put it back into the reclosable aluminium bag. It is not recommended to use the standard solution when the bottle contains less than 10 % of the solution. Therefore, in case of non-transparent bottle, it is important to indicate the amount of the solution used, e.g. on the label. Do not pipette from the bottle. Do not return removed aliquots to bottle

#### Note:

Detailed information about the production, homogeneity, stability, coding, characterization and storing of this CRM are described in the document "Detailed information about the production of aqueous calibration solutions ASTASOL<sup>®</sup>" which is available for download on the website <u>www.analytika.net</u>.

Vorzo cortifikátu	/Varsian of the cartificate: 01	
verze certifikatu	/ Version of the certificate: 01	

2/3

ANALYTIKA<sup>®</sup>, spol. s r.o., Oddělení referenčních materiálů / Department of reference materials Ke Klíčovu 2a/816, CZ-190 00 Praha 9 – Vysočany, Tel/Fax: +420 286 589 616, sales@analytika.net, www.analytika.net

**a analytika** 

#### **Producer:**

ANALYTIKA<sup>®</sup>, spol. s r.o. Department of reference materials Ke Klíčovu 2a/816 190 00 Prague 9 – Vysočany Czech Republic

www.analytika.net sales@analytika.net

Phone/Fax: +420 286 589 616

#### Quality management systems of company ANALYTIKA<sup>®</sup>, spol. s r.o.:

ČSN EN ISO 9001:2016 ČSN EN ISO/IEC 17025:2018 ČSN EN ISO 17034:2017

#### Manager of Department of RM:

Ing. Daniela Weisserová

Date of the first issue of certificate: 09.06.2023 Certificate revision date: Head of production department:

Petránhora

Mgr. Mirka Petránková

Revision of certificate:

Version of certificate: 01

Verze certifikátu / Version of the certificate: 01

3/3

Verze formuláře / Version of the form: 03

ANALYTIKA<sup>®</sup>, spol. s r.o., Oddělení referenčních materiálů / Department of reference materials Ke Klíčovu 2a/816, CZ-190 00 Praha 9 – Vysočany, Tel/Fax: +420 286 589 616, sales@analytika.net, www.analytika.net