

Analytika, spol. s r.o., Oddělení RM, výrobce RM č. 7501, akreditovaný ČIA podle ČSN EN ISO 17034:2017 Analytika, spol. s r.o., Department of RM, Reference Materials Producer No. 7501 accredited by CAI according to ČSN EN ISO 17034:2017

CERTIFICATE

Aqueous calibration solution

ASTASOL® AN90761H

This Certificate is designed in accordance with ISO Guide 31

Category:

Analyte:

Nitrates (NO₃⁻)

Certified reference material

Product code:

AN90761H

NH₄NO₃, 99.999%

Starting primary compound and its purity:

Matrix:

Ultrapure demineralized water (resistivity $\geq 18M\Omega$.cm, 0.22µm filtered)

Density and its expanded uncertainty (k = 2): 0.9994 ± 0.0005 g/cm³ (at 20 °C)

Certified value of concentration and its expanded uncertainty (k = 2) at 20 $^{\circ}C$

$1 \ 000 \pm 2 \ mg/l$

$1\ 001 \pm 3\ mg/kg^*$

*Mass fraction in mg/kg is derived from density

Specification:

Batch No.: 9014

The date of production: 20.06.2023

Shelf life: 5 years from the date of production

The date of first opening of the aluminium bag:

Intended use:

For calibration and validation of analytical methods analysing aqueous solutions such as atomic spectrometry (AAS, AES, ICP-OES, ICP-MS, IC), molecular absorption spectrometry and selected electroanalytical methods

Verze certifikátu / Version of the certificate: 01

1/3

Verze formuláře / Version of the form: 03

ANALYTIKA[®], spol. s r.o., Oddělení referenčních materiálů / Department of reference materials Ke Klíčovu 2a/816, CZ-190 00 Praha 9 – Vysočany, Tel/Fax: +420 286 589 616, sales@analytika.net, www.analytika.net

a analytika

Certification and traceability:

This CRM is certified on the basis of gravimetric preparation. This procedure also ensures a direct traceability to SI unit – kg. Certified concentration, its uncertainty and traceability were verified by gravimetric determination of nitrates and by simultaneous determination of nitrates in independent reference solution (SRM NIST 3185).

Trace impurities in bottled solution (in mg/l):

Determination of trace impurities was performed by ion chromatography. Impurity levels are supplied only for information of the user and should not be used as calibration data.

Bromides	Chlorides	Fluorides	Iodides	Nitrites	Nitrates	Phosphates	Sulfates
< 0.02	< 0.02	< 0.02	< 0.02	< 0.02	А	< 0.02	< 0.05
$< \mathbf{x} - \mathbf{b}$ allow detection limit $\mathbf{A} - \mathbf{a}$							

< x = bellow detection limit A = analyte

Homogeneity and stability:

It has been demonstrated that this CRM is homogeneous and its stability is guaranteed during the whole shelf life provided the solution it kept under conditions presented below.

Storing and instruction for use:

This CRM has to be stored in the original closed bottle between 5 - 30 °C. The producer guarantees a declared shelf life and expiration time provided the CRM is properly stored and professionally handled. The temperature of the solution must be 20 ± 0.5 °C before every use. It is necessary to indicate on this certificate and the label the expiration time, which depends on the date of the first time the aluminium bag was opened. After use, the bottle must be immediately tightly capped, and it is recommended to put it back into the reclosable aluminium bag. It is not recommended to use the standard solution when the bottle contains less than 10 % of the solution. Therefore, in case of non-transparent bottle, it is important to indicate the amount of the solution used, e.g. on the label. Do not pipette from the bottle. Do not return removed aliquots to bottle

Note:

Detailed information about the production, homogeneity, stability, coding, characterization and storing of this CRM are described in the document "Detailed information about the production of aqueous calibration solutions ASTASOL[®]" which is available for download on the website <u>www.analytika.net</u>.

Verze certifikátu / Version of the certificate: 01

2/3

Verze formuláře / Version of the form: 03

ه) analytika

Producer:

ANALYTIKA[®], spol. s r.o. Department of reference materials Ke Klíčovu 2a/816 190 00 Prague 9 – Vysočany Czech Republic

www.analytika.net sales@analytika.net

Phone/Fax: +420 286 589 616

Quality management systems of company ANALYTIKA[®], spol. s r.o.:

ČSN EN ISO 9001:2016 ČSN EN ISO/IEC 17025:2018 ČSN EN ISO 17034:2017

Manager of Department of RM:

Ing. Daniela Weisserová

Date of the first issue of certificate: 20.06.2023

Certificate revision date:

Head of production department:

Petránhora

Mgr. Mirka Petránková

Revision of certificate: Version of certificate: 01

Verze certifikátu / Version of the certificate: 01

3/3

Verze formuláře / Version of the form: 03

ANALYTIKA[®], spol. s r.o., Oddělení referenčních materiálů / Department of reference materials Ke Klíčovu 2a/816, CZ-190 00 Praha 9 – Vysočany, Tel/Fax: +420 286 589 616, sales@analytika.net, www.analytika.net

ه) analytika