ANALYTIKA®, spol. s r.o.

Khodlova 1297/47, CZ-193 00 Praha 9 – Horní Počernice

www.analytika.net



IDENTIFICATION SHEET

ASTASOL® - CH CE281I

Calibration standard solution for chromatography in CERTAN® capillary bottle

This Identification sheet is designed in accordance with ISO Guide 31

VOC MIX-2

Category: Reference material

Product code: CE281I

Analytes:

tetrachloromethane
1,2-dichloroethane
1,1,2-trichloroethylene
1,1,2,2-tetrachloroethylene
trichloromethane

tribromomethane dibromochloromethane bromodichloromethane

Assigned concentration value and expanded uncertainty (k = 2):

 $1000 \pm 10 \,\mu\text{g/ml}$ (each compound, at 20 °C)

| Matrix solvent: |
|---|
| Isooctane |
| Specification: |
| Batch No.: 104 |
| The date of production: 24.01.2023 |
| Shelf life: 24.01.2025 |
| The date of first opening of the bottle: |
| Expiry date: (6 months from the first opening of the bottle within shelf life period, which should be indicated on the label of the bottle as well.) |

Verze IL / Version of the IS: 01

1/2

Verze formuláře / Version of the form: 06

Intended use:

As a calibrator for the determination of the above analytes by chromatographic methods

Gravimetric preparation and traceability:

The single components are weighed by using an electronic balances, dissolved and transferred into a calibrated volumetric flask (class A glassware). The necessary weight of the neat material to compensate the difference of the purity of the compound to 100% is taken into account. The tolerance of the weighing procedure and the dilution error adds to a maximum of \pm 0.5%. Quality control of nominal composition of this solution can be done by comparing the concentrations of this solution with those of an independent internal standard solution.

Packing material:

Amber glass CERTAN® bottle. CERTAN® is a trade-mark of special containers registered by LGC Standards.

Storage:

The bottle should be stored in the dark at temperatures 5 - 10°C. Solution aliquots for analysis should be withdrawn at 20°C immediately after opening the bottle; the bottle should be again tightly capped within 1-2 minutes.

Producer:

ANALYTIKA®, spol. s r.o. Department of reference materials Ke Klíčovu 2a/816 190 00 Prague 9 – Vysočany

www.analytika.net sales@analytika.net

Czech Republic Phone/Fax: +420 286 589 616

Quality management systems of company ANALYTIKA®, spol. s r.o.:

The Producer has a certified quality management system ČSN EN ISO 9001: 2016.

The Producer meets the requirements of ČSN EN ISO 17034:2017: General requirements for the competence of reference material producers.

Head of production department:

Petránhova

Mgr. Mirka Petránková

Date of the first issue of IS: 25.01.2023 Revision of IS:

IS revision date: Version of IS: 01

Verze IL / Version of the IS: 01

2/2

Verze formuláře / Version of the form: 06