

# Analysis of Fatty Acid Methyl Esters (FAME) using Agilent J&W FactorFour VF-WAXms GC Columns

## Application Note


### Author

Helena Jacobse  
Agilent Technologies, Inc.

### Conditions

Technique: GC-FID  
Column: VF-WAXms 15 m x 0.32 mm df =  
0.5  $\mu$ m (part number CP9224)  
Temperature: 200 °C to 250 °C with 20 °C/min  
Carrier Gas: Helium, 0.7 bar

Injector: 250 °C, split 75 mL/min  
Detector: 275 °C, FID  
Sample Size: 0.1  $\mu$ L  
Sample: Rapeseed oil



*Analysis of rapeseed oil using VF-WAXms GC columns*

[www.agilent.com/chem](http://www.agilent.com/chem)

This information is subject to change without notice.

© Agilent Technologies, Inc. 2010

Published in UK, October 08, 2010

SI-00941



**Agilent Technologies**