

Capillary-Grade Hydrocarbon Trap

1/8-inch tube compression fittings—cat.# 21991

1/4-inch tube compression fittings—cat.# 21992



Maximum Operating Pressure: 250 psi

Maximum Operating Temperature: 100 °C

Installation Instructions

1. Remove the protective plug from one end of the trap.
2. Locate nuts and ferrules (two each included) and insert the tubing through the nut and ferrule. Screw the nut onto the fitting until the tubing rests firmly against the fitting shoulder. Tighten finger-tight.
3. Tighten the nut $3/4$ -turn for $1/8$ -inch tubing, and one full turn for $1/4$ -inch tubing.
4. Repeat steps 1 through 3 on the other end of the trap.
5. Pressurize the system and leak check with Restek's Electronic Leak Detector (cat.# 22839). DO NOT use liquid leak detectors.

Health Hazards

Inhalation: May be harmful if inhaled. Material may be irritating to mucous membranes and upper respiratory tract.

Skin Contact: May cause skin irritation.

Eye Contact: May cause eye irritation.

Ingestion: May be harmful if swallowed.

First Aid Measures

Inhalation: Remove to fresh air. If breathing is difficult have a trained individual administer oxygen.

Skin: Wash with soap and water. Get medical attention if irritation develops or persists.

Eye: Flush eyes for 20 minutes. Get medical attention.

Ingestion: Wash out mouth with water, provided person is conscious. Get medical attention.

Note: This document has been prepared based on a component or components of this product. Exposure to the component is not expected through routine laboratory use.

Call Technical Service at 800-356-1688 or 814-353-1300, ext. 4 (or your Restek representative) if you have any questions about this product or any other Restek product.

RESTEK

Chromatography Products
© 2010 Restek Corporation

ISO 9001:2008
REV. 3 FMS0307

#204-08 [005]
Rev. date: 7/10

Restek U.S. • 110 Benner Circle • Bellefonte, PA 16823
phone: 814-353-1300 • 800-356-1688 • fax: 814-353-1309 • www.restek.com

Restek France • phone: +33 (0)1 60 78 32 10 • e-mail: restek@restekfrance.fr

Restek GmbH • phone: +49 (0)6172 2797 0 • e-mail: info@restekgmbh.de

Restek Ireland • phone: +44 (0)2890 814576 • e-mail: restekurope@aol.com

Restek Japan • phone: +81 (3)6459 0025 • e-mail: ryosei.kanaguchi@restek.com

Thames Restek U.K. LTD • phone: +44 (0)1494 563377 • e-mail: sales@thamesrestek.co.uk