



Hydrocarbons, $C_3 - C_4$

Analysis of impurities in 1-butene

Materials Testing & Research

Materials Testing & Research

Authors

Agilent Technologies, Inc.

Introduction

Gas chromatography using an Agilent CP- Al_2O_3/Na_2SO_4 column separates six C_3 to C_4 impurities in 1-butene in 13 minutes.




Agilent Technologies

Conditions

Technique : GC-capillary
Column : Agilent CP-Al₂O₃/Na₂SO₄, 0.32 mm x 50 m fused silica WCOT Al₂O₃/Na₂SO₄ (df = 5 µm) (Part no. CP7565)
Temperature : 110 °C
Carrier Gas : N₂, 110 kPa (1.1 bar, 16 psi)
Injector : Splitter, 20 mL/min
T = 150 °C
Detector : FID, 2 x 10⁻¹² Afs.
T = 200 °C
Sample Size : 100 µL
Concentration Range : 5 - 1000 ppm

Peak identification

1. propane
2. isobutane
3. n-butane
4. trans-2-butene
5. 1-butene
6. isobutene
7. cis-2-butene



www.agilent.com/chem

This information is subject to change without notice.

© Agilent Technologies, Inc. 2011

Printed in the USA

31 October, 2011

First published prior to 11 May, 2010

A00592



Agilent Technologies